

Certificate of Calibration

Fluke Nederland B.V.

Certificate Number:	SA01259494	Date of Calibration:	27 Feb 2024
Receive Condition:	IN TOLERANCE	Date of Recalibration:	27 Feb 2025
Return Condition:	IN TOLERANCE	Place of Calibration:	Eindhoven
Manufacturer:	FLUKE NETWORKS	Temperature within:	(23.0 ± 3) °C
Model:	DSX-5000 INTL	Humidity within:	(45 ± 20) %rh
Serial Number:	3263189		
Description:	1 GHZ DSX CABLE ANALYZER		
Procedure:	Manual Procedure		

Customer: LANCOMAT
PRAHA

Customer Asset ID: -

RMA Number: 606324766

All measurements are traceable to national and/or international standards or have been derived by approved ratio techniques. When possible standards used for this calibration are ISO/IEC 17025 accredited calibrated.

This calibration is performed by a DEKRA certified lab for ISO 9001. This certificate may not be reproduced other than in full. Calibration certificates without signatures, either electronic or handwritten, are not valid.



Issue Date: 27 Feb 2024

Electronically signed

Authorized By

C. Lesanu

Certificate of Calibration

Certificate Number: SA01259494

Remarks

- The calibration status found in this certificate on the top of each results page must be interpreted as:
 - As Found : Data collected before the unit was adjusted and / or repaired
 - As Left : Data collected after the unit has been adjusted and / or repaired
 - Found / Left : Data collected without any adjustment and / or repair performed
- The calibration interval (due date) is the responsibility of the end user.
- According to the European norm 'Operation of electrical installations' NEN-EN 50110-1 release 2013 and the Dutch norm NEN 3140 release 2015 paragraph 5.102.12 through 5.102.16, a safety test is not required. Therefore not performed.
- Temperature conversions (if applicable) are performed according to ISO/IEC 60584:2013 for thermocouples, and ISO/IEC 60751:2022 for resistance temperature devices.

Standards and test-equipment used

Inventory No	Model	Serial No
WP2467	DSX-CALVERST	E000069


Found-Left Report

Model **DSX-5000 CAT 6A/CLASS Fa 1000MHz Copper Module**
 Serial Number **3263189**

Test date 27-Feb-24
 Page 1 of 3

NEXT


NEXT Artifact SN 3386461



Pass Worst margin: 0.400 at 1 MHz in pair 36-12. Worst accuracy at each frequency shown.

CDNEXT


CDNEXT Artifact SN 3386449



Pass Worst margin: 0.330 at 31.13 MHz in pair 45-78. Worst accuracy at each frequency shown.

CMRL

CMDMRL Artifact SN 3386324



Pass Worst margin: 0.390 at 366 MHz in pair 12. Worst accuracy at each frequency shown.

■ Measured difference of DSX and reference laboratory equipment added to measurement accuracy of reference laboratory equipment. Worst accuracy at each frequency shown.

■ Corresponding measurement accuracy requirement for nominally compliant Level IV or Level 2G/VI field tester.

Found-Left Report


Model **DSX-5000 CAT 6A/CLASS Fa 1000MHz Copper Module**
 Serial Number **3263189**

Test date 27-Feb-24

Page 2 of 3

RL


CMDMRL Artifact SN 3386324



Pass Worst margin: 0.390 at 915 MHz in pair 45. Worst accuracy at each frequency shown.

TCL


TCL Artifact SN 3386457



Pass Worst margin: 0.500 at 3.5 MHz in pair 45. Worst accuracy at each frequency shown.

IL

ILFEXT Artifact SN 3386158



Pass Worst margin: 0.660 at 1.88 MHz in pair 12. Worst accuracy at each frequency shown.

- Measured difference of DSX and reference laboratory equipment added to measurement accuracy of reference laboratory equipment. Worst accuracy at each frequency shown.
- Corresponding measurement accuracy requirement for nominally compliant Level IV or Level 2G/VI field tester.

Found-Left Report


Model **DSX-5000 CAT 6A/CLASS Fa 1000MHz Copper Module**
 Serial Number **3263189**

Test date 27-Feb-24

FEXT

Page 3 of 3

ILFEXT Artifact SN 3386158



Pass Worst margin: 0.410 at 71 MHz in pair 12-36. Worst accuracy at each frequency shown.

■ Measured difference of DSX and reference laboratory equipment added to measurement accuracy of reference laboratory equipment. Worst accuracy at each frequency shown.

■ Corresponding measurement accuracy requirement for nominally compliant Level IV or Level 2G/VI field tester.

Loop Resistance

Loop Resistance Artifact SN 3386276

	Measured	Expected	Limit	
Resistance on pair 12	0.16	0.00	0.80	Pass
Resistance on pair 36	49.99	49.80	0.60	Pass
Resistance on pair 45	99.93	99.80	1.60	Pass
Resistance on pair 78	452.94	453.00	4.00	Pass

Resistance imbalance

Resistance Unbalance Artifact SN 3386085

	Measured	Expected	Limit	
Resistance on pair 12	0.23	0.00	0.80	Pass
Resistance on pair 36	25.07	24.90	0.90	Pass
Resistance on pair 45	12.27	12.13	0.90	Pass
Resistance on pair 78	24.25	24.05	0.90	Pass
Resistance imbalance on pair 12	0.00	0.00	0.05	Pass
Resistance imbalance on pair 36	0.01	0.00	0.13	Pass
Resistance imbalance on pair 45	0.33	0.32	0.06	Pass
Resistance imbalance on pair 78	0.85	0.85	0.12	Pass

DSX-8000 only: M_IL and M_FEXT measurements validate the ability of the DSX-8000 to make measurements with DSX-5000 adapters.

M IL Not applicable M_ILFEXT Artifact SN -

M FEXT Not applicable M_ILFEXT Artifact SN -